

# NICKEL, URINE

## Orderable – NIU / NIU24

Turnaround Time: 10 days

### Specimen:

24-hour urine collected in new 24-hour urine container or random urine

### Collection Information:

#### Trace Elements Collection Requirements

### Reference Ranges:

#### SI Units:

Random Urine:  
0-10 µmol/mol creatinine

24 Hour Urine:  
0.0-85.2 nmol/d

#### Conventional Units:

Random Urine:  
0-5.2 µg/g creatinine

24 Hour Urine:  
0.0-5.0 µg/d

Reference Ranges are based on Non-Occupationally exposed population.

### Interpretive Comments:

Find Interpretive Comment and Clinical Information [here](#):

### Specimen Stability:

Ambient: 14 days  
Refrigerated: 11 months  
Frozen: 11 months



**Laboratory:**  
Trace Elements Lab



**Requisition:**  
TRACE ELEMENTS  
REQUISITION



**Method of Analysis:**  
ICP-MS/MS



**Test Schedule:**  
Batched analysis