

BLD - Blood Leak Detector

Steps to rule out microscopic blood when this occurs:

1. Send an effluent sample in mauve tube
2. Order "CDF" (cell count and differential, fluid) in PowerChart
3. Under specimen type choose "miscellaneous"
4. Under body site choose "see comment"
5. Under label comment put "effluent sample query RBC's present?"
6. Call Lab and let them know you are sending the sample and it needs to be done STAT and why